Bassett Healthcare Network Point of Care Eye pH QC Result Log

		Location:						
оН 5.0 LOT#:	Exp.Date:	pH 7.0 LOT#:	Exp.Date:	_ pH Paper LOT#:	_Exp.Date:			

Date	Time	pH 7.0 QC Result Acceptable 6.5-7.5	pH 5.0 QC Result Acceptable 4.5-5.5	Initials of Operator	Patient entered into Epic POC 120 Initials	POC Review

NOTE: All data is reviewed by the Point of Care technical staff during regularly scheduled site visits. The review encompasses information from the last review date to the current review date. Staff performing tests are responsible to ensure that results are within acceptable range.